We thank the Reviewers for their professional comments on the manuscript. The manuscript has been revised according to the Reviewers' comments and suggestions. Each Reviewer comment and its corresponding response are outlined below, including a sample of manuscript text where the modifications were applied, as well as a reference to page and line numbers in the revised manuscript. All modifications to the manuscript are also tracked in the document using Microsoft Word's "track changes" feature, where deletions and additions are marked with red strikeout, and red underlined text, respectively.

Referee: 1

Page 2, line 2: Change "portion" to "proportion".

Authors' response:

Thank you for pointing this out, "proportion" is indeed a more suitable term in this context. The corrected sentence in the manuscript now reads:

Page 3, line 2: "Osteoporosis contributes to a large *proportion* of fractures worldwide."

Referee: 1

Page 2, line 4: Is "patients" the right word? Do you really mean "the at-risk population"?

Authors' response:

The authors acknowledge that the term "patients" refers to people who are already admitted into medical care, and might be misleading in the context of preventive screening. Your recommendation is good, and the sentence has been revised as follows:

Page 3, lines 4-5: "Therefore, organisations such as the World Health Organisation recommend screening of the *at-risk population* before fractures occur [1]."

Referee: 1

Page 2, line 11: Delete "successful". I do not believe that everyone agrees that the technique really is successful – if it was, there would be no reason to try to improve it.

Authors' response:

Thank you for this comment. Indeed, the correlations between quantitative ultrasound parameters and gold-standard DXA bone mineral density measurements are rather modest. However, in this context, the authors wanted to emphasise the viability of ultrasonic techniques for fracture prediction, as discussed in Glüer, 2008. We chose to use the term "successful" to emphasise the potential diagnostic value that calcaneal through-transmission ultrasound measurements in terms of predicting fractures before they occur. However, we agree with the Reviewer and the sentence has been modified to read as follows:

Page 3, line 12: "Calcaneal through-transmission ultrasound measurement has been established for the determination of quantitative ultrasound (QUS) parameters of relevant diagnostic value [5]–[11]."

Referee: 1

Page 2, line 14: I do not think that the determinants really are significant. The paper (you are the authors) report that one of the correlation coefficients is only 0.75.

Authors' response:

This is an excellent comment. The choice of wording may be slightly overambitious on the authors' part. The word "significant" has now been removed from the sentence, which now reads as follows:

Page 3, line 15: "Moreover, quantitative parameters related to sound scattering derived from pulse-echo (PE) ultrasound measurements *can* serve as determinants of bone structure and mechanical properties, in addition to being more feasible in its application at more commonly fractured skeletal sites [12]–[14]."

Referee: 1

Page 2, line 18: This is in vitro, not in vivo.

Authors' response:

Thank you for pointing this out, this sentence was accidentally incorrectly referenced. The correct reference is Karjalainen 2008, which extended the preceding DFUS study (reference [16]) to an *in-vivo* setting. The reference [15] has been corrected in the revised manuscript as follows:

[15] J. Karjalainen, J. Töyräs, T. Rikkinen, J. S. Jurvelin, and O. Riekkinen, "Dual-frequency ultrasound technique minimizes errors induced by soft tissue in ultrasound bone densitometry," *Acta Radiol.*, vol. 49, no. 9, pp. 1038–41, Nov. 2008.

Referee: 1

Page 2, line 20: Reference [16] is a key paper, but it reports on the results of in vitro experiments.

Authors' response:

Thank you for this comment, this sentence is lacking the same citation to Karjalainen as above. The reference has been corrected in the revised manuscript.

Referee: 1

Page 3, line 21: You state later (page 4 line 4) that the transducer operated at 5 MHz. Delete "100 MHz",

Authors' response:

Thank you for pointing this out, the authors acknowledge that the description of the measurement devices is not clear enough. In this context, the 100MHz is meant to refer to the sampling frequency of the receiver's data acquisition electronics. The sentence has been clarified in the manuscript as follows:

Page 4, line 22: "...a 32-channel phased-array multiplexer together with a 100 MHz *sampling rate* ultrasound pulser-receiver and an 8-bit 100 MHz digital-analogue converter combination..."

Referee: 1

Page 4, line 4: What was the bandwidth of the "5 MHz" transducer?

Authors' response:

Indeed, this information was accidentally omitted from the manuscript. We measured a centre frequency of 4.49 MHz and a -6dB bandwidth of 3.22 – 5.27 MHz in degassed distilled water. The revised manuscript now reads:

Page 5, line 5: "A 32-element 5 MHz (mean frequency: 4.49 MHz, 3.22-5.27MHz, -6 dB) linear phased-array ultrasound transducer (Optel ultrasonic technology ltd.) with an active aperture of 20 mm and element pitch of 0.625 mm was custom-built for the purposes of this experiment."

Referee: 1

Page 4, line 13: What was the measured attenuation coefficient of the oil, and what was its frequency dependence?

Authors' response:

The attenuation frequency dependence can be described using the following fit: $\alpha = 0.06865f^{1.726}$, where α and f are attenuation coefficient in dB/cm and frequency in MHz, respectively. This information has been added to the revised manuscript at the end of the materials and methods section as follows:

Page 7, line 20: "In our measurements, we determined the average attenuation coefficient and speed of sound of rapeseed oil to be 1.1 dB/cm at 5MHz and 1436.1 m/s, respectively."

Referee: 1

Page 4, line 16: The term "interfering" (here and elsewhere; sometimes you use the term "interference" seems to me to be inappropriate. It implies that the phase is important. I should prefer "tissue-simulating".

Authors' response:

Thank you for pointing this out. The authors agree with this comment, and the sections of the text containing term "interfering layer" have been modified to favour the term "tissue-simulating" instead. These modifications are applied throughout the revised manuscript in the following locations:

Page 4, line 18: "As *tissue-simulating* materials of variable thickness over the object (polished steel plate) we used rapeseed oil and water."

Page 5, lines 18-19: "To evaluate the ability of the phased-array probe to correctly determine the thickness of the *two tissue-simulating layers*, we used a syringe to accurately lay down a layer of rapeseed oil over the layer of degassed water."

Page 5, line 21: "Total *tissue-simulating* layer thickness was varied from 40 mm to 80 mm in 10 mm increments"...

Page 7, lines 8-9: "The *tissue-simulating* layer thicknesses can be calculated with the DFUS technique [16], [18] according to the following equations:"...

Page 7, line 15: ..."where x , c , α , Δt and $A(f)$ are the thicknesses of the *tissue-simulating* layer, speed of sound, attenuation coefficient, time of flight and ultrasound pressure amplitude spectrum, respectively."

Page 8, lines 2-4: "In each measurement setup, the total *tissue-simulating* layer thicknesses, as calculated using DFUS, were linearly correlated with the known *tissue-simulating* layer thicknesses, regardless of the focusing settings"

Page 8, lines 6-7: "DFUS underestimated the total thickness of *tissue-simulating* layers when the thickness was less than"...

Page 8, line 12: "In all setups, the DFUS technique could accurately determine the compositions of the two *tissue-simulating* layers only when the focusing was optimal (figure 3)."

Page 8, line 16: "Under non-optimal focus, the absolute error in DFUS-determined *tissue-simulating* layer thickness increased (figure 3)."

Page 9, lines 1-2: "The results suggest that the thickness of the *tissue-simulating* layers can be correctly determined only when the ultrasound beam is optimally focused on the surface of the object."

Page 9, line 6: "The focusing settings had little effect on the accuracy with which the total thickness of the two *tissue-simulating* layers could be determined (figure 2a)."

Page 9, line 9: ..."significantly decreased ability of the system to accurately determine the compositions of *tissue-simulating* layers (figure 3)."

Page 10, lines 6-7: "This laboratory experiment shows that phased-arrays, when combined with DFUS calculations, have ability to correctly estimate thicknesses of *tissue-simulating* layers"...

Referee: 1

Page 5, line 2-3: I think that you should explain why the signal has its maximum amplitude.

Authors' response:

Thank you for bringing this up, the authors agree that the sentence requires some elaboration. The max amplitude arises from having a superposition of constructive interference from all 32 incoming wave fronts. This is now explained in the revised manuscript as follows:

Page 6, lines 4-6: "Thus, the 40th focal point was perfectly located at the surface of the steel disc in each measurement, so that all wave fronts arrive at the point at the same time, producing higher wave pressure through constructive interference, resulting in maximum amplitude in the signal arising from the reflection at that point."

Referee: 1

Page 5, line 17: What was the speed of sound in the oil?

Authors' response:

Our measurements reported an average speed of sound of 1436.1 ± 4.0 m/s across all measurements in which oil was involved. This information has now been added to the revised manuscript at the end of the materials and methods section on page 7:

Page 7, line 20: "In our measurements, we determined the average attenuation coefficient and speed of sound of rapeseed oil to be 1.1 dB/cm at 5MHz and 1436.1 m/s, respectively."

Referee: 1

Page 6, line 1-2: To what extent did this reduce the signal amplitude? What was the significance of the non-directionality of the receiver?

Authors' response:

Thank you for your comments. Ultrasound angle of incidence is an important component in maintaining a high signal-to-noise ratio, which is why optimal incidence was maintained throughout all measurements. Since the receiving element was located at the centre of the transducer, the ultrasound pressure wave always travelled the shortest possible distance from the sample to the receiver, thus we always recorded the highest possible signal amplitude. Obviously, recording using only one transducer decreases the total amplitude of received ultrasound signal. Limiting receiving to the centre transducer was done in order to simplify the calculations by eliminating the need for complicated phase-correction for multiple signals arriving at different phases and times. We considered the decrease in received amplitude a minor problem compared to the advantages achieved through this method.

Referee: 1

Page 7, line 18-19: How could normal incidence be ensured in vivo?

Authors' response:

This is an excellent point, and is indeed a difficult one to address. The benefit of phased-array transducers is that the individual elements of the linear array are spread out across the total width of the transducer, which enables the conduction of a series of A-line scans in rapid succession. This reveals the orientation of the bone within the tissue, which enables the correct alignment of the transducer for conducting more accurate focused beam ultrasound measurements.

This dialogue has been outlined in the revised manuscript in the discussion section on page 10 as follows:

Page 10, lines 8-10: "*However, linear-array B-mode imaging prior the focusing of ultrasound beam to the bone surface ensures optimal ultrasound angle of incidence, thus further improving signal-to-noise ratio.*"

Referee: 1

Page 8, line 2-4: Why do you think that this is the "most likely" reason?

Authors' response:

Thank you for this comment. The frequency spectrum and overall quality of arriving ultrasound signals is a critical component in the DFUS calculation. Non-optimal focusing settings cause unknown amounts of both constructive and destructive interference in the ultrasound pressure waves, which translates into a scrambled echo signal arising from the surface of the sample. This can be observed from the raw RF signal data, which is used for FFT spectrum analysis and ultimately the DFUS calculation. A similar observation was made *in-silico* and *in-vitro* by Malo et al. in 2011. (Reference [17] in the manuscript)
This sentence has now been clarified in the revised manuscript as follows:

Page 9, lines 9-11: "This is congruent with findings in previous *in-silico* and *in-vitro* studies, and is likely due to interference and phase cancellation of the multiple wave fronts upon striking the sample surface, resulting in distortion of the frequency spectrum of the echo signal [17]."

Referee: 1

Page 9, line 2-3: What "further investigations" are required?

Authors' response:

This is an interesting comment. The next logical step is to apply the techniques described to an *in-vivo* setting with a large sample size. This would involve reference Computer Tomography measurements of bone density as well as Magnetic Resonance Imaging of soft tissue composition so that the viability of the Ultrasound technique for diagnostic purposes could be determined.
This has now been elaborated in the revised manuscript as follows:

Page 10, lines 12-16: "However, further investigations are desired to determine if the methodology can be extended into *in vivo* applications. Such experiments should involve the application of the measurement technique to a large population, with reference bone density and soft tissue composition measurements using Computer Tomography and Magnetic Resonance Imaging for final validation."

Referee: 2

This paper is well written and presented and obviously follows on from previous work done by the group. This is a laboratory based experiment and i was wondering how the authors are planning to use the results in an *in vivo* setting.

Authors' response:

Thank you for your comment. This work is a necessary step towards the clinical *in-vivo* application of the Dual Frequency Ultrasound technique. This paper validates the viability of using the DFUS technique in conjunction with phased array ultrasound transducers in order to overcome some of the issues relating to lens-focused transducers with fixed focal length, as discussed in previous papers. Our next step is to apply this technique *in-vivo* on a large population and contrasting it to reference measurements using Computer Tomography and Magnetic Resonance Imaging.

Author's response to Reviewer's comments: TMET-2015-0128

Hans Linder, Markus K H Malo, Jukka Liukkonen, Jukka S Jurvelin and Juha Töyräs:

"Phased-array ultrasound technology enhances accuracy of dual frequency ultrasound measurements – towards improved ultrasound bone diagnostics"

This dialogue has been included in the revised manuscript as follows:

Page 10, lines 12-16: "However, further investigations are desired to determine if the methodology can be extended into *in vivo* applications. Such experiments should involve the application of the measurement technique to a large population, with reference bone density and soft tissue composition measurements using Computer Tomography and Magnetic Resonance Imaging for final validation."

References:

[Glüer, 2008] - IEEE transactions on ultrasonics, ferroelectrics, and frequency control, vol. 55, no. 7, july 2008

[Karjalainen, 2008] - J. Karjalainen, J. Töyräs, T. Rikkinen, J.S. Jurvelin & O. Riekkinen (2008) Dual-Frequency Ultrasound Technique Minimizes Errors Induced by Soft Tissue in Ultrasound Bone Densitometry, Acta Radiologica, 49:9, 1038-1041