

## Supporting Information

### Research Article

#### **Sensitive and accurate determination of 168 micropollutants including pharmaceuticals and pesticides in surface water and wastewater samples with direct injection using jet stream ESI LC-MS/MS**

Oltan Canlı<sup>a\*</sup>, Elmas Oktem Olgun<sup>a</sup>, Barış Güzel<sup>a</sup>, Muammer Kaplan<sup>b</sup>

---

<sup>a</sup> Environment and Cleaner Production Institute, TUBITAK Marmara Research Center, 41470 Gebze, Kocaeli, Turkey.

<sup>b</sup> Chemical Technology Institute, TUBITAK Marmara Research Center, 41470 Gebze, Kocaeli, Turkey.

\* Corresponding Author e-mail: oltan.canli@tubitak.gov.tr, phone: +90-262-677-29-22; fax: +90-262-641-23-09).

Co-Author's e-mail and phone: ((elmas.olgün@tubitak.gov.tr, phone: +90-262-677-29-97), (guzelbaris08@gmail.com, phone: +90-262-677-42-62), (muammer.kaplan@tubitak.gov.tr, +90-262-677-38-42))

---

**Table S1.** Inter-laboratory test results for pesticides in wastewater

<b>Compound Name</b>	<b>Reported Result (<math>\mu\text{g/L}</math>)</b>	<b>Satisfactory Range (<math>\mu\text{g/L}</math>)</b>	<b>z-score</b>
Carbaryl	60	41.1-69.1	1.17
Carbofuran	93	58-110	0.675
Diuron	97.7	44.6-147	0.0762
Methomyl	165	64.1-175	1.64
Propham	53	20.2-74.5	0.417
Bentazone	3.1	2.28-4.51	1.32
Picloram	4	1.96-7.42	-0.504
Azinphos methyl	18.2	5.21-21.0	1.28
Diazinon	3	1.29-3.64	0.915
Dichlorvos	<2	<2	-
Dimethoate	14	6.08-18.2	0.621
Malathion	10	4.75-10.4	1.73