A Low Complex Algorithm for Detection of Sleep Apnea from ECG Signals using Sliding SSA Combined with PCA, KPCA, and SPPCA

Umesh Kumar Naik M, Muhammad Zubair, *Student Member, IEEE*, Rajesh Kumar Tripathy, *Member, IEEE* and Ramanjaneyulu Kongara, *Member, IEEE*.

Abstract—Sleep apnea is a potentially life-threatening sleep condition in which breathing stops and resumes repeatedly. It is caused by breathing pauses during sleep, which leads to frequent awakenings. As we all know, computational time and efficiency are essential in the healthcare industry; to address this issue, we proposed an algorithm that performs more computations in less time without compromising the machine learning model's performance. This study employs a unique technique called Sliding Singular Spectrum Analysis (SSSA) to decompose and de-noise the ECG signals. To identify the significant apnea and non-apnea components from the pre-processed ECG data and to reduce the dimensionality, we used Principal Component Analysis (PCA), Kernel PCA (KPCA), and Sub-Pattern based PCA (SPPCA). These characteristics were then used to train and evaluate various machine learning models, including KNN, SVM, GaussianNB, SGD, and XGBoost, to distinguish between apnea and non-apnea ECG data. The publicly available Physionet Apnea-ECG database is used for the simulation of the proposed algorithm. To verify the performance of machine learning models, we have calculated various metrics like accuracy, precision, recall and F1 score. The validation of the proposed method is done by comparing the classification metrics with the latest state-of-the-art works.

Index Terms—Sleep Apnea, Sliding Singular Spectrum Analysis, Dimensionality Reduction, Kernel-PCA, PCA, Sub-Pattern based PCA

1 INTRODUCTION

SLEEP apnea is among the most prevalent sleep disorders, characterised by breathing pauses while sleeping, also termed as apneic events, which cause frequent awakenings [1]. Obstructive Sleep Apnea (OSA) occurs when the airway is obstructed by the throat muscles, Central Sleep Apnea (CSA) occurs when the signals that govern breathing are disrupted, and hypopnea occurs when breathing becomes short. Hypopnea can be classified as either obstructive or central hypopnea. Although some studies have shown that around 49.7% of male and 23.4% of female people have sleep-disordered breath conditions [2], many cases remain undetected, as patients never know their illness. These people are predisposed to hypertension, cardiac arrhythmias, heart attacks, and strokes [3] [4]. Some studies also suggest that people with sleep apnea are more likely to engage in collisions with the motor vehicle [5].

Polysomnography (PSG), a sleep test conducted in a hospital or at home that requires the assessment and supervision of a physician, is now used to evaluate sleep-related respiratory problems. Despite the fact that PSG is

the most essential diagnostic tool, its high price and lack of comfort limit its diagnostic value. Because the number of beds available for PSG recording and the number of qualified sleep specialists available for analysis are both restricted, wait periods can be exorbitant. In the United Kingdom, these wait periods range from 2 to 10 months, whereas in the United States, they range from 7 to 60 months [6]. Furthermore, there is a lot of intra- and inter-scorer variability. As a result, the development of technologies that are more portable and less invasive is critical [7] [8] [9].

Despite the significant time and money invested on manual sleep staging, there are still flaws. Generally, the agreement of two distinct ratings is unacceptable [10]- [11] [12] [13] [14] [15]. The inter-rater reliability (IRR) between two scorers using the current sleep scoring criteria is typically about 0.78 [10], as assessed by Cohen's kappa. However, due to poor N1 sleep scoring, the dependability amongst worldwide sleep centres can be as low as 0.58 to 0.63 [11] [12] [13]. The agreement of N1 amongst sleep laboratories in Europe is only about 0.46, and it might be as low as 0.19 to 0.31 between international centres. Furthermore, if a person has a medical condition, the general reliability of manual sleep staging may be reduced even further. For example, the dependability of OSA patients is lower than that of healthy people [14] [15]. Automatic scoring approaches have the potential to enhance sleep staging uniformity across hospitals and healthcare systems. Furthermore, automated techniques capable of precise sleep staging with a small number of observed signals might make the measuring process easier to follow and save money.

- Umesh Kumar Naik M, Muhammad Zubair, Dr. Ramanjaneyulu Kongara are with the Department of Electronics and Communication Engineering, Prasad V. Potluri Siddhartha Institute of Technology, Vijayawada, AP, India (e-mail: umeshkumar.mudavath; muhammadzub17; kongara.raman)@gmail.com
- Dr. Rajesh Kumar Tripathy is with the Department of Electronics and Electrical Engineering, Birla Institute of Technology (BITS), Hyderabad, Telangana, India (e-mail: tripathyrk@hyderabad.bits-pilani.ac.in)

Manuscript received April 19, 2021; revised August 26, 2021.

As a result, we present a thorough benchmark of several classifiers and four chosen features based on a single ECG signal. Five alternative classifiers for detecting sleep apnea from ECG data were examined in accordance with this. Furthermore, these classifiers were put to the test in three separate dimensionality reduction scenarios, each with its own set of characteristics (also extracted from the signal). In comparison to PSG's traditional screening schemes, the proposed approaches might provide practitioners with a reliable, easy, and cost-effective diagnostic tool.

2 LITERATURE REVIEW

Sleep apnea can be diagnosed from ECG and EEG signals. In [16], an algorithm for the automatic detection of sleep apnea from single-lead ECG was developed using two features derived from the ECG; namely the standard deviation and the serial correlation coefficients of the RR interval time series along with LS-SVM classifier, which achieved an accuracy of 0.84% on Physionet and 0.83% on Leuven datasets. In [17], an automatic apnea detection algorithm is proposed using single lead EEG Signal based on Rician Modeling of Feature Variation in Multi-band EEG Signal along with KNN Classifier and achieved an accuracy of 0.91% on Physionet and 0.86 on MIT-BIH datasets. In [18], they proposed a wearable, accurate, and energy efficient system for monitoring obstructive sleep apnea on a long-term basis. based on monitoring the patient using a single-channel electrocardiogram achieved an accuracy of 0.88% on Physionet data.

Other techniques based on machine learning, which have a better learning capacity, can automatically discover more complex structures and generate more accurate predictions. Commonly used methods include SVM [19], Logistic Regression (LR) [20], KNN [19], [20], Linear Discriminant Analysis (LDA) [20], [21], Gaussian Processes (GP) [22] and Artificial Neural Networks (ANN) [23]. The classifying rules or characteristics of each sleep stage were not clearly stated in these systems. Previous research, on the other hand, have often depended on extensive pre-processing, such as converting signals into 2D pictures that contain spectral information or limiting signals to a small number of predetermined characteristics. Furthermore, deep learning models based on healthy people's study datasets have usually lost accuracy when applied to populations with sleep disorders like OSA.

In [24] the following classifiers were built using a single-lead ECG signal: kNN, Multilayer Perceptron Neural Network (MLPNN), and SVM, Least-Square Support Vector Machine (LS-SVM). The RBF kernel-based LS-SVM produced better accuracy, according to the experimental data. Similarly, in [25] and [26] Adaptive boosting (AdaBoost) and SVM were developed to identify normal and apnea episodes, respectively, based on a single-lead ECG. To detect sleep apnea episodes, the authors in [27] retrieved the EDR signal from a single-lead ECG and used the following classifiers: ANN, SVM, kNN, Linear Discriminant (LD), and Quadratic Discriminant (QD). As per their reports, the ANN with two hidden layers gives better performance. Similarly, in their proposal for sleep apnea diagnosis [28], the authors utilised the HRV signal and applied the following classifiers:

ANN, BN, kNN, and SVM. The linear kernel SVM achieved the best results. Furthermore, the authors pointed out that the effectiveness of a feature extraction technique varies depending on the categorization method. In [29] with edge computing ideas in mind, the authors suggested a sleep apnea detection system. The authors used RF, Extremely Randomized Trees, SVM, NB, AdaBoost, kNN, and LR to assess the system's performance using data from a single-channel ECG sensor. Despite the limited amount of features available, the SVM combined with RBF kernel was shown to have the best accuracy. The authors of [30] suggested a sleep apnea detection sensor based on a micro electro-mechanical system (MEMS). The primary objective was to track diaphragm motions during breathing exercise. The suggested model's classifier was the artificial neural network (ANN). In addition, in [31], the authors presented a wearable for ambulatory sleep apnea monitoring. To differentiate between normal and apnea episodes, the model employed a single-lead ECG and an SVM classifier. In contrast, authors suggested a method based on the nasal airflow signal in [32]. The SVM was chosen as the classifier for evaluating the system's performance.

The following deep learning models were constructed using the ECG data in [33]: DNN, 1D-CNN, 2D-CNN, Recurrent Neural Networks (RNN), Long Short-Term Memory (LSTM), and Gated-Recurrent Unit (GRU). In addition, the authors recommended combining time series signals with either the 1D CNN or the GRU. In [34], on the other hand, authors utilised a CNN model to analyse a single-channel nasal pressure signal. Furthermore, [35] used LSTM-RNN to integrate the IHR and SpO2 (LSTM-RNN).

Computational time and efficiency are essential in the healthcare industry. Here, we proposed an algorithm that performs more computations in less time without compromising the performance of machine learning models. Usually, reducing the number of variables in a data set reduces accuracy; nevertheless, the answer to dimensionality reduction algorithms (DRA) is to trade little accuracy for simplicity. Because smaller data sets are simpler to examine and visualise, and because ML models can analyse data more easily and quickly without having to deal with unnecessary factors. To summarise, our proposed DRA's purpose is to reduce the number of variables in a data set while retaining as much information as possible.

The purpose of this study is to provide an automated approach for detecting sleep apnea utilising the ECG signal and sliding singular spectrum analysis (SSSA). The SSSA decomposition of non-stationary signals into reconstructed components (RCs) or modes is a variation of the singular spectrum analysis (SSA) method. Mono-component separation from simulated and actual non-stationary signals has been shown using this approach. Because it is a data-driven approach, the SSSA delivers efficient results for multi-scale analysis of non-stationary signals without making any assumptions about the basis functions.

3 PROPOSED METHODOLOGY

The proposed method mainly comprises of five steps, (i) Pre-processing of ECG signals using sliding SSA, (ii) Extraction of four attribute's from sliding SSA output such

as dominant frequency, energy, spectral entropy, and spike rhythmicity, (iii) Calculating the principal components of apneic and non-apneic ECG signals using PCA, KPCA, and SPPCA on the four attributes individually, (iv) Classification of apneic and non-apneic ECG signals using various machine learning models (KNN, SVM, GaussianNB, SGD, and XGBoost), and (v) Comparison of computational time and performance measures (accuracy, precision, recall and F1 Score) between SSSA-BR, SSSA-PCA, SSSA-KPCA, and SSSA-SPPCA. The hierarchy of proposed sliding SSA combined with KPCA, PCA and SPPCA algorithm are shown in Fig. 3.

3.1 Data description

This study made use of the apnea-ECG database, which is a public database available at Physionet [36]. This database comprises 70 ECG datasets, each of which lasts around 8 hours. Each ECG dataset has a 100 Hz sample frequency. For the patients in this database, the Apnea-Hypopnea-Index (AHI) ranges from 0 to 83. An expert assessed 35 of the 70 ECG recordings as apnea (labelled as 1) or normal (labelled as 0) minute by minute. Furthermore, based on the number of apnea minutes and AHI, each recording is categorised as "apnea," "borderline," and "normal." We have considered 35 single lead ECG recordings from this database. In this work, we have used each one minute segment of the ECG signal which consist of 6000 samples. The one minute segment of apneic and non-apneic ECG signal is shown in Fig. 1. And Singular spectrum's of apnea and non-apnea episodes extracted using SSSA (for $L=50$) is shown in Fig. 2.

3.2 Sliding singular spectrum analysis (SSSA)

The S-SSA can be derived from the standard SSA algorithm [37]. The SSA algorithm is a mixture of two important steps known as: (i) Decomposition and (ii) Reconstruction. The apnea or non-apnea ECG signal (Y_n) is usually corrupted with other artifacts. To remove these unwanted sources from the ECG signal, we perform decomposition at the first stage and the reconstruction of the noise-free ECG signal (X_n) is performed at the second stage.

Fig. 1. One minute segment of apneic and non-apneic ECG signal

3.2.1 Decomposition

The SSA decomposition stage is broken down into two parts: embedding and Singular Value Decomposition (SVD). Let us consider an ECG data matrix $\mathbf{E}$ of size $(K \times M)$ as mentioned in (1), where $(K = N - L + 1)$, N and L reflect the number of samples and the length of the window, respectively.

$$\mathbf{E} = \begin{pmatrix} s_1 & s_2 & \dots & s_K \\ s_2 & s_3 & \dots & s_{K+1} \\ \vdots & \vdots & \ddots & \vdots \\ s_L & s_{L+1} & \dots & s_N \end{pmatrix} \quad (1)$$

Secondly, SVD is performed on the matrix $\mathbf{E}$. The implementation of SVD on $\mathbf{E}$ can be represented as:

$$\mathbf{E} = \mathbf{V}\mathbf{D}\mathbf{U}^T \quad (2)$$

where, $\mathbf{V}$ and $\mathbf{U}$ are the unitary matrices and $\mathbf{D}$ is a diagonal matrix.

Most of the time, SVD is used to find the symmetrical eigen vectors. $\mathbf{C} = \mathbf{E}\mathbf{E}^T$ gives the covariance matrix $\mathbf{C}$ for the information matrix $\mathbf{E}$, which corresponds to the eigenvalues and eigenvectors as $\lambda_1, \lambda_2, \dots, \lambda_M$ and $\mathbf{V}_1, \mathbf{V}_2, \dots, \mathbf{V}_M$ individually. The i_{th} component of $\mathbf{U}$ can be expressed as in (3),

$$\mathbf{u}_i = \frac{\mathbf{E}^T \mathbf{v}_i}{\sqrt{\lambda_i}} \quad (3)$$

where, $i = 1, 2, \dots, M$. The trajectory matrix can therefore be represented as follows,

$$\mathbf{E} = \sum_{i=1}^M \mathbf{E}_i = \sum_{i=1}^M \sqrt{\lambda_i} \mathbf{v}_i \mathbf{u}_i^T \quad (4)$$

where trajectory matrix $\mathbf{E}_i$ can be defined as

$$\mathbf{E}_i = \mathbf{v}_i \mathbf{v}_i^T \mathbf{E} \quad (5)$$

The term $\mathbf{v}_i \mathbf{v}_i^T$ in (5) forms a subspace, which is formed by the eigenvector $\mathbf{v}_i$.

Fig. 2. Singular spectrum's of apnea and non-apnea episodes extracted using SSSA (for $L=50$)

Fig. 3. Hierarchy of proposed sliding SSA combined with KPCA, PCA and SPPCA algorithms

3.2.2 Reconstruction

Constructing the SSA again comprises of two steps, which are mainly grouping and diagonal averaging. It has been observed that to construct again each elementary component for any value of L , the usual SSA fails. So in this paper we apply clustering after rebuilding each elementary component, and for reconstruction we have used the Hankelization process.

$$\mathbf{E}_i = [e(k, p)] = s_i(n) \quad (6)$$

where, k and p are the indices of the elements of $\mathbf{E}_i$ and $s_i(n)$ is the i_{th} component of $\mathbf{E}_i$.

To initialize this algorithm, we allocate each reconstructed elementary component to a unique class. Then to build a novel class the algorithm looks for the closest two classes and puts them together. And we keep on doing this process until the desired number of classes or the maximum dissimilitude between two classes is achieved. For comparing the classes C_i and C_j which have a lot of time series data, the minimum distance between the two distinct time series of each class gives the dissimilarity.

$$D(C_i, C_j) = \min_{x \in C_i, y \in C_j} d(x, y) \quad (7)$$

the distance function d is: $R^N \times R^N \rightarrow [0, 1]$

$$d(x, y) = 1 - \frac{|\langle x, y \rangle|}{\|x\| \|y\|} \quad (8)$$

where, $\langle x, y \rangle = x^T y$ and $\|x\| = \sqrt{\langle x, x \rangle}$

The *autoSSA* function is used to perform the hierarchical clustering on the $s_i(n)$ components which are, $[s_1(n), s_2(n), s_3(n), \dots, s_R(n)]$. Finally, the j_{th} cluster components are assigned as indices of J_j where, $j = 1, 2, 3, \dots, r$.

In the last step, the components within each cluster are designated as the j_{th} reconstructed components [38].

3.3 Calculating attributes on SSSA output

3.3.1 Dominant frequency

DF and SE, two frequency domain characteristics, have been explored. To get frequency domain characteristics, first convert the time domain signal into the frequency domain using DFT as described in (9).

$$S(k) = \sum_{n=0}^{N-1} s[n] e^{-j \frac{2\pi}{N} kn} \quad (9)$$

The maximum peak frequency of each sub-domain is known as dominant frequency. This characteristic is particularly beneficial in the identification of sleep apnea since it yields superior outcomes. For an ECG signal, we computed the dominant frequency of each epoch of size 32.

3.3.2 Spectral entropy

For each sub-pattern, spectral entropy is computed between two frequencies, f_1 and f_2 . It may be written as follows: (10).

$$SE = \sum_{f_i=f_1}^{f_2} P_n(f_i) \times \log \left(\frac{1}{P_n(f_i)} \right) \quad (10)$$

Where, f_i is the frequency component ranges from f_1 to f_2 , and $P_n(f_i)$ is the normalized power spectrum component at f_i .

After that, by dividing the SE by $\log(N)$, the normalised spectral entropy (NSE) is calculated. In this work, SE of an ECG signal is calculated using 32-point FFT.

Fig. 4. Spectral Entropy before dimensionality reduction

Fig. 6. Spectral Entropy after applying SPPCA

Fig. 5. Spectral Entropy after applying PCA

Fig. 7. Spectral Entropy after applying KPCA

$$NSE = \frac{SE}{\log(N)} \quad (11)$$

3.3.3 Energy of the signal

Summing the square values of each sub-pattern sample as shown in (11) may be used to compute the energy of each sub-pattern.

$$Energy(E_i) = \sum_{j=1}^N D_{ij}^2 \quad (12)$$

where, D can be number of samples in each sub-pattern.

3.4 Dimensionality reduction of the attributes

3.4.1 PCA

PCA is most commonly used dimensionality reduction algorithm to compress larger dimensions into smaller ones while keeping the majority of the original dataset's information, which is crucial when classifying the Attributes.

The following approach is used to apply PCA on data of dimension $p \times q$ (where p is the number of data and q is the number of dimensions of each element):

- To begin, the extracted Attributes are used to build a covariance matrix of size $(p \times p)$.
- Calculate the covariance matrix's eigenvalues (λ^i) and associated eigenvectors E^i .
- Choose the greatest eigenvalues and eigenvectors for 'k' ($k < m$) eigenvalues. Let E^i represent the set of these 'k' chosen eigenvectors.
- To obtain the PCs for attribute vectors, projection of attribute vectors S^i onto E^i is carried out in order to generate attributes with a reduced dimension according to the following relationship stated in (13).

$$P_i = S_i \times E_i \quad (13)$$

Fig. 4 depicts the Spectral Entropy of ECG signal before dimensionality reduction and Fig. 5 depicts the output ECG signal after dimensionality reduction using PCA. We employed two PCA variations to minimise the dimension-

Fig. 8. Computational time of spectral entropy for various machine learning models

ality of attributes in this study. The following are the two algorithms:

3.4.2 SPPCA

The signal is initially split into several equal sub-parts in SPPCA. We separated an ECG signal into eight parts because they produced the desired results. PCA is applied to each sub-part of the signal in the second phase, and a k number of principal components (PCs) is chosen. The k should be smaller than the signal's dimension in this case. All of the sub-parts are horizontally concatenated after the dimensionality reduction to provide adequate global data. Finally, the classifier is fed with the reduced dimension signal [39]. Fig. 6 depicts the output ECG signal after dimensionality reduction using SPPCA.

3.4.3 KPCA

KPCA is a technique for enhancing the characteristics of ECG data. It also offers a variety of kernel selection algorithms. Gaussian Kernel was employed in this case. KPCA works on the premise of mapping data to a higher-dimensional component space F with a nonlinear relationship to the input space [40]. Because KPCA is a circle and PCA is a straight line for the biggest difference between the projections of the points onto the eigenvector (new coordinates), KPCA has greater variance than PCA. Fig. 7 depicts the output ECG signal after dimensionality reduction using KPCA.

KPCA is a PCA method that accounts for non-linearities in ECG data by performing a non-linear mapping of the data onto a higher dimensional space using a kernel function, after which PCA is performed.

In this method, if X is the input space, and ψ is a non-linear transformation, the kernel $\mathbf{K}$ is constructed from their inner product, i.e., $\mathbf{K} = k(x_p, x_q) = \psi(x_p)^T \psi(x_q)$. The eigen-delineation of the centered kernel matrix is given in (14)

$$\mathbf{K} = \mathbf{H}\mathbf{K}\mathbf{H} = \mathbf{U}\mathbf{\Lambda}\mathbf{U}^T \quad (14)$$

where, $\mathbf{H} = \mathbf{I} - 1/N\mathbf{1}\mathbf{1}^T$ is the centering matrix, $\mathbf{I}$ is the $q \times q$ identity matrix, $\mathbf{1} = [1, 1, \dots, 1]^T$ is a $q \times 1$ vector, $\mathbf{U} = [w_1, \dots, w_q]$ with $w_i = [w_1, \dots, w_q]$, where, $\mathbf{T}$ is the

matrix containing the eigenvectors and $\mathbf{\Lambda} = \text{diag}(\lambda_1, \dots, \lambda_q)$ contains the corresponding eigenvalues. Finding the eigenvectors and the eigenvalues, the reconstructed signal is calculated by a procedure similar to that used in the PCA method.

3.5 Time complexity reduction using PCA, SPPCA and KPCA

Consider the model parameters of class P with dimension d as $P_1, P_2, P_3, \dots, P_N$. The temporal complexity may be calculated using PCA as follows:

$$T(PCA) = O(Nd^2 + d^3) \quad (15)$$

The input patterns are split into L sub-patterns in the proposed SPPCA algorithm. As a result, the temporal complexity of SPPCA is:

$$T(SPPCA) = O(L(Nr^2 + r^3)) \quad (16)$$

4 EXPERIMENTAL RESULTS

Initially, from 1000000 we took 900000 samples of apnea-ECG data, and segmentation of the ECG signal is performed using SSSA. The original signal was divided into 150×6000 channels by setting the window length ($L = 40$). After segmentation, we have extracted four attributes namely Dominant Frequency (DF), Energy (EN), Spectral Entropy (SE), and Spike Rhythmicity (SR). Now, these four attributes dimensions are reduced using DRA in order to decrease the computational time of machine learning models. We used three DRA's to extract the various principal components from the chosen attributes. The output of these three DRA's are now fed to machine learning models for classification of given signal into either apnea or non-apnea classes. The proposed method is performed on Apple MacBook powered with an M1 chip. We used *autoSSA* function on MATLAB 2021a to implement SSSA-BR, SSSA-PCA, SSSA-KPCA, and SSSA-SPPCA algorithm. The steps taken for implementing the DRA's is as follows: -

- After downloading the data from the Physionet Dataset, the apnea-ECG dataset of patient a01.mat (of length 900000) is imported into MATLAB.
- The single signal is decomposed into multivariate signal using SSSA
- Dominant Frequency (DF), Energy (EN), Spectral Entropy (SE), and Spike Rhythmicity (SR) attributes are computed on SSSA's output.
- SSSA-PCA, SSSA-KPCA, and SSSA-SPPCA are performed on the attributes in order to reduce the dimensions of the features.
- Size of data after performing DRA's on every attribute is as follows: -
  - SSSA-Before Reduction : 300×6000
  - SSSA-PCA : 300×100
  - SSSA-KPCA : 300×100
  - SSSA-SPPCA : 300×160
- And these attributes are saved in a .csv delimited text file.

TABLE 1
Comparison of different classification parameters for SSSA-BR and SSSA-KPCA algorithms using various machine learning models

Features	SSSA-BR					SSSA-KPCA			
	Models	Accuracy	Precision	Recall	F1 score	Accuracy	Precision	Recall	F1 score
Energy	KNN	0.99	1.00	1.00	1.00	0.51	0.24	0.50	0.33
	SVM	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	GaussianNB	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	SGD	1.00	1.00	1.00	1.00	0.71	1.00	1.00	1.00
	XGBoost	0.99	1.00	1.00	1.00	1.00	1.00	1.00	1.00
Dominant Frequency	KNN	0.99	1.00	1.00	1.00	0.51	0.24	0.50	0.33
	SVM	0.99	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	GaussianNB	0.99	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	SGD	1.00	1.00	1.00	1.00	0.57	1.00	1.00	1.00
	XGBoost	0.99	1.00	1.00	1.00	1.00	1.00	1.00	1.00
Spectral Entropy	KNN	0.99	1.00	1.00	1.00	0.48	0.24	0.50	0.33
	SVM	0.98	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	GaussianNB	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	SGD	1.00	1.00	1.00	1.00	0.54	1.00	1.00	1.00
	XGBoost	0.99	1.00	1.00	1.00	1.00	1.00	1.00	1.00
Spike Rhythmicity	KNN	0.99	1.00	1.00	1.00	0.51	0.24	0.50	0.33
	SVM	0.98	1.00	1.00	1.00	0.97	0.97	0.97	0.97
	GaussianNB	0.99	1.00	1.00	1.00	1.00	0.97	0.97	0.97
	SGD	1.00	1.00	1.00	1.00	0.54	0.97	0.97	0.97
	XGBoost	0.98	1.00	1.00	1.00	1.00	0.97	0.97	0.97

TABLE 2
Comparison of different classification parameters for SSSA-PCA and SSSA-SPPCA algorithms using various machine learning models

Features	SSSA-PCA					SSSA-SPPCA			
	Models	Accuracy	Precision	Recall	F1 score	Accuracy	Precision	Recall	F1 score
Energy	KNN	1.00	0.95	0.94	0.94	0.96	0.92	0.92	0.92
	SVM	0.97	0.97	0.97	0.97	0.96	0.97	0.97	0.97
	GaussianNB	1.00	0.97	0.97	0.97	1.00	0.97	0.97	0.97
	SGD	1.00	0.97	0.97	0.97	1.00	0.97	0.97	0.97
	XGBoost	1.00	0.97	0.97	0.97	1.00	0.97	0.97	0.97
Dominant Frequency	KNN	1.00	0.95	0.94	0.94	0.88	0.89	0.87	0.86
	SVM	0.97	0.97	0.97	0.97	1.00	1.00	1.00	1.00
	GaussianNB	1.00	0.97	0.97	0.97	1.00	1.00	1.00	1.00
	SGD	1.00	0.97	0.97	0.97	0.98	1.00	1.00	1.00
	XGBoost	1.00	0.97	0.97	0.97	1.00	1.00	1.00	1.00
Spectral Entropy	KNN	1.00	0.91	0.88	0.88	1.00	1.00	1.00	1.00
	SVM	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	GaussianNB	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	SGD	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	XGBoost	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
Spike Rhythmicity	KNN	1.00	0.97	0.97	0.97	0.93	0.89	0.87	0.86
	SVM	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	GaussianNB	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	SGD	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
	XGBoost	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00

- Finally, the attributes are fed into ML models for classification.

To identify ECG data as apnea or non-apnea, we used a variety of machine learning algorithms such as KNN, SVM, GaussianNB, SGD, and XGBoost. The steps taken for implementing the models is as follows:

- The dataset was annotated and then it was shuffled so that each data point contributes independently and there is no biasing.
- The attribute extracted in the previous selection are fed into ML models for classification.
- The classification data was divided in the ratio of 8:2, i.e, 80% for training and 20% for testing.
- The above mentioned models were then trained and their performance metrics were calculated using the testing data.

For gradient boosting, we took number of estimators as 100, learning rate as 1.0, max depth as 1 and random state as a 0. Similarly, for AdaBoost classifier we have taken number of estimators as 100 and random state as 0. We have considered RBF kernel to perform SVM. For XGBosst the number of 100 estimators are considered. Finally, we have calculated the computational time of all the classifiers for SSSA-BR, SSSA-KPCA, SSSA-PCA, and SSSA-SPPCA algorithms which is presented in Table 3. The classification was done using Python 3.8.5 on a PC loaded with 4GB of RAM and powered with intel pentium processor. Performance metrics (such as accuracy, precision, recall, and F1 score) for SSSA-BR, and SSSA-KPCA are shown in Table 1. Performance metrics for SSSA-PCA, and SSSA-SPPCA are shown in Table 2. Fig. 8 depicts the computational time of spectral entropy for various machine learning models.

TABLE 3
Comparison of Computational time (in Seconds) for SSSA-BR, SSSA-KPCA, SSSA-PCA, and SSSA-SPPCA algorithms using various machine learning models

Features	Models	SSSA-BR	SSSA-KPCA	SSSA-PCA	SSSA-SPPCA
Dominant Frequency	KNN	118.08	0.06	0.07	0.13
	SVM	1.15	0.03	0.03	0.05
	GaussianNB	0.08	0.02	0.03	0.03
	SGD	0.08	0.03	0.02	0.02
	XGB	3.9	0.13	0.13	0.37
Energy	KNN	102.78	0.06	0.07	0.13
	SVM	1.33	0.05	0.03	0.06
	GaussianNB	0.09	0.03	0.03	0.04
	SGD	0.08	0.02	0.02	0.02
	XGB	4.33	0.13	0.3	0.22
Spectral Entropy	KNN	116.99	0.07	0.09	0.11
	SVM	0.3	0.04	0.05	0.05
	GaussianNB	0.08	0.04	0.02	0.03
	SGD	0.09	0.02	0.02	0.02
	XGB	3.97	0.13	0.14	0.24
Spike Rythmicity	KNN	107.35	0.07	0.08	0.11
	SVM	1.99	0.04	0.05	0.04
	GaussianNB	0.08	0.02	0.04	0.03
	SGD	0.07	0.01	0.01	0.01
	XGB	4.07	0.01	0.14	0.22

TABLE 4
Comparison of our work with other state-of-the art implementations

Author	Method	Accuracy (%)
Qi Shen et al. [41]	Multiscale dilation 1D-CNN	89.40
Nuno Pombo et al. [42]	HRV and EDR signals	82.12
Woon Jung et al. [43]	A new predictor algorithm for AHI prediction	94.50
Carolina et al. [44]	LS-SVM with RBF kernel	85.00
Himali Singh et al. [45]	Sliding SSA with stacked autoencoder (SAE-DNN)	94.30
R.K. Tripathy et al. [46]	fast and adaptive empirical mode decomposition (FAEMD) with SVM	78.67
Sokolovsky et al. [47]	Deep Convolutional Neural Networks	81.00
Asghar Zarei et al. [48]	sequential forward feature selection (SFFS) with random forest	93.90
P. Janbakhshi et al. [49]	ECG variance and phase space reconstruction area	98.40
R.K. Tripathy et al. [50]	Intrinsic band functions (IBFs) with kernel extreme learning machine (KELM)	86.80
Proposed Work	Sliding SSA-Before Reduction (BR)	99.40
	Sliding SSA-Kernal PCA (KPCA)	99.60
	Sliding SSA-Principal Component Analysis (PCA)	100
	Sliding SSA-Sub-Pattern based PCA (SPPCA)	98.60

5 DISCUSSION

Several automated sleep apnea diagnosis techniques based on ML classification have recently been developed to replace traditional approaches. As shown in Table IV, the performance of our suggested technique was compared to that of existing methodologies for the classification of apnea and non-apnea classes from the ECG signal. In the recent study, Qi Shen et al. [41] has developed an algorithm to detect Sleep Apnea. In this work OSA detection method based on a multiscale dilation attention 1-DCNN (MSDA-1DCNN) and a weighted-loss time-dependent (WLTD) classification model were employed and obtained an accuracy of 89.4%

respectively. Similarly, Nuno Pombo et al. [42] have used Sgolay filter to extract the Heart Rate Variability (HRV) and the ECG-Derived Respiration (EDR) and obtained an accuracy of 82.12%.

Moreover, Da Woon Jung et al. [43] has developed an a new predictor of AHI Prediction Using ECG Acquired During the Sleep-Onset Period while they obtained an accuracy of 94.5%. Similarly, Carolina Varon et al. [44] have also developed an algorithm using two features in heart rate variability analysis; The first feature uses the principal components of the QRS complexes and The second feature extracts the information shared between respiration and heart rate using orthogonal subspace projections. And achieved an accuracy of 85% using least-squares SVM classifier, using an RBF kernel. In another study, Himali Singh et al. [45] have developed an algorithm using sliding mode SSA combined with classifiers, such as the stacked autoencoder based deep neural network (SAE-DNN), and SVM. They obtained accuracy values of 94.3%, and 72% respectively.

Moreover, R.K.Tripathy et al. [46] have developed an automated sleep apnea detection using using bivariate

TABLE 5
Average of accuracy for all the features

Features	SSSA-BR	SSSA-KPCA	SSSA-PCA	SSSA-SPPCA
DF	0.992	0.996	0.992	0.990
EN	0.816	0.844	0.804	0.804
SE	0.994	0.994	1.000	1.000
SR	0.972	0.984	1.000	0.986

fast and adaptive empirical mode decomposition (FAEMD) combined with SVM and the random forest (RF) classifiers achieved an average accuracy of 78.67% and using the 10-fold cross-validation method achieved an average accuracy of 73.13% respectively. Moreover, Michael Sokolovsky et al. [47] have designed a DCNN architecture for automated sleep stage classification of human sleep EEG and EOG signals and the accuracy obtained by this method was 81%. Similarly, AsgharZarei et al. [48] have developed an algorithm using sequential forward feature selection (SFFS) technique and fed into the random forest for classification and achieved an average accuracy of 93.9% in per-segment classification. Moreover, P.Janbakhshi et al. [49] have designed an algorithm using ECG variance and phase space reconstruction area. The accuracy obtained by this method was 100% in subject-based apnea detection on independent test data. Similarly, R.K.Tripathy [50] have developed automated detection of sleep apnea using by extraction of features from the intrinsic band functions (IBFs) of both EDR and HRV signals, and the classification using kernel extreme learning machine (KELM) and achieved an accuracy of 74.64% respectively.

More focus has been placed in the proposed study on developing an ML framework uses the time-frequency domain characteristics acquired from the ECG data. Fig. 9 depicts the Average accuracy plot for all four features. To minimise the apnea feature dimensions, we proposed three DRAs: SSSA-PCA, SSSA-KPCA, and SSSA-SPPCA. Furthermore, we employed five different machine learning models to validate the efficiency of the proposed framework, which produced better results when compared to [41] - [50] and had a 100% accuracy rate. In Table III, all of the classification parameters are compared in detail. The main objective of our work is to reduce the computational time while classification of the ECG signal with less complexity and high accuracy. which reduces the burden on the neurologist to perform evaluation of the test results. This complexity is gradually decreased by employing suggested DRA; which shorten the time it takes to detect seizures and reduce human error while examining ECG signals.

Fig. 9. Average accuracy plot for all four features

6 CONCLUSION

In this paper, we proposed an DRA's combined with SSSA that performs more computations in less time without compromising the ML model's performance. Here, more emphasis is given to reduce computational time and give better efficiency as they are very essential in the health-care. We combined sliding SSA with PCA, SPPCA, and KPCA to identify the apnea attributes from the ECG signals. Initially, the ECG signals were decomposed using sliding SSA technique. Then, the Principal components of decomposed ECG signal were calculated using PCA, SPPCA and KPCA. Finally, the apnea and non-apnea dimension reduced attributes were given as an input to five ML classifiers. The classification metrics of all the models are shown in Table I, and Table II. And compared the computational time of ML classifiers using SSSA-BR, SSSA-PCA, SSSA-KPCA, and SSSA-SPPCA as shown in Table I and Table II. The proposed algorithm achieved an highest accuracy of 100% on XGBoost classifier. To validate the proposed algorithm, we have compared our work with the latest state-of-the-art works as shown in Table IV. Average of accuracy for all the features are shown in Table V. In future, we are planning to detect sleep apnea in ECG signals using deep learning algorithms based on multivariate features extracted using convolutional autoencoders.

REFERENCES

- [1] C. Guilleminault, A. Tilkian and W. C. Dement, "The sleep apnea syndromes", *Annu. Rev. Med.*, vol. 27, no. 1, pp. 465-484, 1976.
- [2] R. Heinzer et al., "Prevalence of sleep-disordered breathing in the general population: The hypnolaus study", *Lancet Respiratory Med.*, vol. 3, no. 4, pp. 310-318, 2015.
- [3] V. K. Somers et al., "Sleep apnea and cardiovascular disease", *Circulation*, vol. 118, no. 10, pp. 1080-1111, 2008.
- [4] H. K. Yaggi, J. Concato, W. N. Kernan, J. H. Lichtman, L. M. Brass and V. Mohsenin, "Obstructive sleep apnea as a risk factor for stroke and death", *New England J. Med.*, vol. 353, no. 19, pp. 2034-2041, 2005.
- [5] A. Sassani, L. J. Findley, M. Kryger, E. Goldlust, C. George and T. M. Davidson, "Reducing motor-vehicle collisions costs and fatalities by treating obstructive sleep apnea syndrome", *Sleep*, vol. 27, no. 3, pp. 453-458, 2004.
- [6] W. W. Flemons, N. J. Douglas, S. T. Kuna, D. O. Rodenstein and J. Wheatley, "Access to diagnosis and treatment of patients with suspected sleep apnea", *Amer. J. Respiratory Crit. Care Med.*, vol. 169, no. 6, pp. 668-672, 2004.
- [7] D. Bliwise, N. G. Bliwise, H. C. Kraemer and W. Dement, "Measurement error in visually scored electrophysiological data: Respiration during sleep", *J. Neuroscience Methods*, vol. 12, no. 1, pp. 49-56, 1984.
- [8] N. A. Collop, "Scoring variability between polysomnography technologists in different sleep laboratories", *Sleep Med.*, vol. 3, no. 1, pp. 43-47, 2002.
- [9] S. T. Kuna et al., "Agreement in computer-assisted manual scoring of polysomnograms across sleep centers", *Sleep*, vol. 36, no. 4, pp. 583-589, 2013.
- [10] H. Danker-Hopfe et al., "Interrater reliability for sleep scoring according to the Rechtschaffen & Kales and the new AASM standard", *J. Sleep Res.*, vol. 18, no. 1, pp. 74-84, 2009.
- [11] U. J. Magalang et al., "Agreement in the scoring of respiratory events and sleep among international sleep centers", *Sleep*, vol. 36, no. 4, pp. 591-596, 2013.
- [12] X. Zhang et al., "Process and outcome for international reliability in sleep scoring", *Sleep Breath.*, vol. 19, no. 1, pp. 191-195, 2015.
- [13] R. S. Rosenberg and S. Van Hout, "The american academy of sleep medicine inter-scorer reliability program: Respiratory events", *J. Clin. Sleep Med.*, vol. 10, no. 4, pp. 447-454, 2014.

- [14] T. Penzel, X. Zhang and I. Fietze, "Inter-scorer reliability between sleep centers can teach us what to improve in the scoring rules", *J. Clin. Sleep Med.*, vol. 9, no. 1, pp. 89-91, 2013.
- [15] R. G. Norman, I. Pal, C. Stewart, J. A. Walsleben and D. M. Rapoport, "Interobserver agreement among sleep scorers from different centers in a large dataset", *Sleep*, vol. 23, no. 7, pp. 901-908, 2000.
- [16] C. Varon, A. Caicedo, D. Testelmans, B. Buyse and S. Van Huel, "A Novel Algorithm for the Automatic Detection of Sleep Apnea From Single-Lead ECG," in *IEEE Transactions on Biomedical Engineering*, vol. 62, no. 9, pp. 2269-2278, Sept. 2015, doi: 10.1109/TBME.2015.2422378.
- [17] A. Bhattacharjee, S. Saha, S. A. Fattah, W. -P. Zhu and M. O. Ahmad, "Sleep Apnea Detection Based on Rician Modeling of Feature Variation in Multi-band EEG Signal," in *IEEE Journal of Biomedical and Health Informatics*, vol. 23, no. 3, pp. 1066-1074, May 2019, doi: 10.1109/JBHI.2018.2845303.
- [18] G. Surrel, A. Aminifar, F. Rincon, S. Murali and D. Atienza, "Online Obstructive Sleep Apnea Detection on Medical Wearable Sensors," in *IEEE Transactions on Biomedical Circuits and Systems*, vol. 12, no. 4, pp. 762-773, Aug. 2018, doi: 10.1109/TBCAS.2018.2824659.
- [19] B. Yilmaz, M. H. Asyali, E. Arkan, S. Yetkin and F. Özgen, "Sleep stage and obstructive apneic epoch classification using single-lead ECG", *Biomed. Eng. Online*, vol. 9, no. 1, pp. 1-14, 2010.
- [20] J. V. Marcos, R. Hornero, D. Álvarez, F. del Campo and C. Zamarrón, "Assessment of four statistical pattern recognition techniques to assist in obstructive sleep apnoea diagnosis from nocturnal oximetry", *Med. Eng. Phys.*, vol. 31, no. 8, pp. 971-978, 2009.
- [21] P. De Chazal, C. Heneghan, E. Sheridan, R. Reilly, P. Nolan and M. O'Malley, "Automatic classification of sleep apnea epochs using the electrocardiogram", *Proc. IEEE Comput. Cardiology*, pp. 745-748, 2000.
- [22] S. Gutta, Q. Cheng, H. Nguyen and B. Benjamin, "Cardiorespiratory model-based data-driven approach for sleep apnea detection", *IEEE J. Biomed. Health Inform.*, vol. 22, no. 4, pp. 1036-1045, Jul. 2018.
- [23] P. Várady, T. Micsik, S. Benedek and Z. Benyó, "A novel method for the detection of apnea and hypopnea events in respiration signals", *IEEE Trans. Biomed. Eng.*, vol. 49, no. 9, pp. 936-942, Sep. 2002.
- [24] H. Sharma and K. K. Sharma, "An algorithm for sleep apnea detection from single-lead ECG using Hermite basis functions", *Comput. Biol. Med.*, vol. 77, pp. 116-124, Oct. 2016.
- [25] A. R. Hassan, "Computer-aided obstructive sleep apnea detection using normal inverse Gaussian parameters and adaptive boosting", *Biomed. Signal Process. Control*, vol. 29, pp. 22-30, Aug. 2016.
- [26] B. Fatimah, P. Singh, A. Singhal and R. B. Pachori, "Detection of apnea events from ECG segments using Fourier decomposition method", *Biomed. Signal Process. Control*, vol. 61, Aug. 2020.
- [27] P. Janbakhshi and M. B. Shamsollahi, "Sleep apnea detection from single-lead ECG using features based on ECG-derived respiration (EDR) signals", *IRBM*, vol. 39, no. 3, pp. 206-218, Jun. 2018.
- [28] B. Sulistyo, N. Surantha and S. M. Isa, "Sleep apnea identification using HRV features of ECG signals", *Int. J. Electr. Comput. Eng.*, vol. 8, no. 5, pp. 3940, Oct. 2018.
- [29] A. Stojanovski, E. Zdravetski, S. Koceski and V. Trajkovik, "Real-time sleep apnea detection with one-channel ecg based on edge computing paradigm", *ICT Innov.*, vol. 2018, pp. 124-134, Sep. 2018.
- [30] A. H. Yüzer, H. Sümbül, M. Nour and K. Polat, "A different sleep apnea classification system with neural network based on the acceleration signals", *Appl. Acoust.*, vol. 163, Jun. 2020.
- [31] G. Surrel, A. Aminifar, F. Rincon, S. Murali and D. Atienza, "Online obstructive sleep apnea detection on medical wearable sensors", *IEEE Trans. Biomed. Circuits Syst.*, vol. 12, no. 4, pp. 762-773, Aug. 2018.
- [32] B. L. Koley and D. Dey, "Automatic detection of sleep apnea and hypopnea events from single channel measurement of respiration signal employing ensemble binary SVM classifiers", *Measurement*, vol. 46, no. 7, pp. 2082-2092, Aug. 2013.
- [33] U. Erdenebayar, Y. J. Kim, J.-U. Park, E. Y. Joo and K.-J. Lee, "Deep learning approaches for automatic detection of sleep apnea events from an electrocardiogram", *Comput. Methods Programs Biomed.*, vol. 180, Oct. 2019.
- [34] S. H. Choi, H. Yoon, H. S. Kim, H. B. Kim, H. B. Kwon, S. M. Oh, et al., "Real-time apnea-hypopnea event detection during sleep by convolutional neural networks", *Comput. Biol. Med.*, vol. 100, pp. 123-131, Sep. 2018.
- [35] R. K. Pathinarupothi, D. P. J., E. S. Rangan, G. E. A., V. R. and K. P. Soman, "Single sensor techniques for sleep apnea diagnosis using deep learning", *Proc. IEEE Int. Conf. Healthcare Informat. (ICHI)*, pp. 524-529, Aug. 2017.
- [36] T. Penzel, G.B. Moody, R.G. Mark, A.L. Goldberger, J.H. Peter The apnea-ECG database *Computers in Cardiology 2000*, vol. 27, IEEE (2000), pp. 255-258 (Cat. 00CH37163)
- [37] M. Zubair, U. Kumar Naik M and R. Ahamed Shaik, "EMG Artifacts Removal from Multi-Channel EEG Signals using Multi-Channel Singular Spectrum Analysis," 2020 IEEE Applied Signal Processing Conference (ASPCON), 2020, pp. 183-187, doi: 10.1109/ASPCON49795.2020.9276684.
- [38] J. Harmouche, D. Fourer, F. Auger, P. Borgnat and P. Flandrin, "The Sliding Singular Spectrum Analysis: A Data-Driven Non-stationary Signal Decomposition Tool," in *IEEE Transactions on Signal Processing*, vol. 66, no. 1, pp. 251-263, 1 Jan.1, 2018, doi: 10.1109/TSP.2017.2752720.
- [39] M. Zubair et al., "Detection of Epileptic Seizures from EEG Signals by Combining Dimensionality Reduction Algorithms with Machine Learning Models," in *IEEE Sensors Journal*, doi: 10.1109/JSEN.2021.3077578.
- [40] M. M. V. Gualsaquí, E. I. P. Vizcaino, M. J. Flores-Calero and E. V. Carrera, "ECG signal denoising through kernel principal components," 2017 IEEE XXIV International Conference on Electronics, Electrical Engineering and Computing (INTERCON), 2017, pp. 1-4, doi: 10.1109/INTERCON.2017.8079670.
- [41] Q. Shen, H. Qin, K. Wei and G. Liu, "Multiscale Deep Neural Network for Obstructive Sleep Apnea Detection Using RR Interval From Single-Lead ECG Signal," in *IEEE Transactions on Instrumentation and Measurement*, vol. 70, pp. 1-13, 2021, Art no. 2506913, doi: 10.1109/TIM.2021.3062414.
- [42] N. Pombo, B. M. C. Silva, A. M. Pinho and N. Garcia, "Classifier Precision Analysis for Sleep Apnea Detection Using ECG Signals," in *IEEE Access*, vol. 8, pp. 200477-200485, 2020, doi: 10.1109/ACCESS.2020.3036024.
- [43] D. W. Jung, S. H. Hwang, Y. J. Lee, D. Jeong and K. S. Park, "Apnea-Hypopnea Index Prediction Using Electrocardiogram Acquired During the Sleep-Onset Period," in *IEEE Transactions on Biomedical Engineering*, vol. 64, no. 2, pp. 295-301, Feb. 2017, doi: 10.1109/TBME.2016.2554138.
- [44] C. Varon, A. Caicedo, D. Testelmans, B. Buyse and S. Van Huffel, "A Novel Algorithm for the Automatic Detection of Sleep Apnea From Single-Lead ECG," in *IEEE Transactions on Biomedical Engineering*, vol. 62, no. 9, pp. 2269-2278, Sept. 2015, doi: 10.1109/TBME.2015.2422378.
- [45] Himali Singh, Rajesh Kumar Tripathy, Ram Bilas Pachori, Detection of sleep apnea from heart beat interval and ECG derived respiration signals using sliding mode singular spectrum analysis, *Digital Signal Processing*, Volume 104, 2020, 102796, ISSN 1051-2004, <https://doi.org/10.1016/j.dsp.2020.102796>.
- [46] R.K. Tripathy, Pranjali Gajbhiye, U. Rajendra Acharya, Automated sleep apnea detection from cardio-pulmonary signal using bivariate fast and adaptive EMD coupled with cross time-frequency analysis, *Computers in Biology and Medicine*, Volume 120, 2020, 103769, ISSN 0010-4825, <https://doi.org/10.1016/j.combiomed.2020.103769>.
- [47] M. Sokolovsky, F. Guerrero, S. Paisarnsrisomsuk, C. Ruiz and S. A. Alvarez, "Deep Learning for Automated Feature Discovery and Classification of Sleep Stages," in *IEEE/ACM Transactions on Computational Biology and Bioinformatics*, vol. 17, no. 6, pp. 1835-1845, 1 Nov.-Dec. 2020, doi: 10.1109/TCBB.2019.2912955.
- [48] Asghar Zarei, Babak Mohammadzadeh Asl, Performance evaluation of the spectral autocorrelation function and autoregressive models for automated sleep apnea detection using single-lead ECG signal, *Computer Methods and Programs in Biomedicine*, Volume 195, 2020, 105626, ISSN 0169-2607, <https://doi.org/10.1016/j.cmpb.2020.105626>.
- [49] P. Janbakhshi, M.B. Shamsollahi, Sleep Apnea Detection from Single-Lead ECG Using Features Based on ECG-Derived Respiration (EDR) Signals, *IRBM*, Volume 39, Issue 3, 2018, Pages 206-218, ISSN 1959-0318, <https://doi.org/10.1016/j.irbm.2018.03.002>.
- [50] R.K. Tripathy, Application of intrinsic band function technique for automated detection of sleep apnea using HRV and EDR signals, *Biocybernetics and Biomedical Engineering*,

Volume 38, Issue 1, 2018, Pages 136-144, ISSN 0208-5216,
<https://doi.org/10.1016/j.bbe.2017.11.003>.

Umesh Kumar Naik M has completed his bachelor's degree in Electronics and Communication Engineering at Prasad V. Potluri Siddhartha Institute of Technology, Vijayawada, India; affiliated to JNTU Kakinada, Andhra Pradesh, India.

He has published three research papers in various IEEE conferences and one paper in IEEE Sensors Journal. He completed two research internships at IIT Guwahati, Assam, India. He is presently an active student member

at IETE organisation. His research interests include Biomedical Signal processing, Time-Frequency Analysis, Biomedical Image processing, Machine learning, Deep learning, and Brain Computer Interface for healthcare.

Ramanjaneyulu Kongara (Member, IEEE) received his PhD from Andhra University in 2012. He is working as a Professor in the Department of ECE, PVP Siddhartha Institute of Technology, Vijayawada. His research interests are in the areas of Digital Watermarking, Image Processing and its Applications, Artificial Intelligence and Wireless communications.

Muhammad Zubair (Student Member, IEEE) has completed his bachelor's degree in Electronics and Communication Engineering at Prasad V. Potluri Siddhartha Institute of Technology, Vijayawada, India; affiliated to JNTU Kakinada, Andhra Pradesh, India.

He has published four research papers in various IEEE international conferences and two papers in IEEE Sensors Journal. He completed three research internships at IIT Guwahati, Assam, India. He is presently an active student

member at IEEE and IETE organisations. His research interests include Biomedical Signal and Image processing, Time-Frequency Analysis, Machine Learning, Deep Learning and Brain Computer Interface.

Rajesh Kumar Tripathy (Member, IEEE) received the B.Tech degree in Electronics and Telecommunication Engineering from the Biju Patnaik University of Technology (BPUT), Odisha, India, in 2009, and the M.Tech degree in Biomedical Engineering from the National Institute of Technology (NIT) Rourkela, Rourkela, India, in 2013, and the Ph.D. degree in Machine Learning for Biomedical signal processing under Electronics and Electrical Engineering (EEE) department from the Indian Institute of Technology

(IIT) Guwahati, India in 2017. He worked as an Assistant Professor at the faculty of engineering and technology (FET), Siksha 'O' Anusandhan Deemed to be University from March 2017 to June 2018. From July 2018, he is working as an Assistant Professor in the Department of Electrical and Electronics Engineering (EEE), Birla Institute of Technology and Science (BITS), Pilani, Hyderabad Campus. His research interests are Machine learning, Deep Learning, Biomedical Signal Processing, Sensor Data Processing, Medical Image Processing, and the Internet of Things (IoT) for healthcare. He has published research papers in reputed international journals and conferences. He has served as a reviewer for more than 15 scientific journals and served as a technical program committee (TPC) member in various national and international conferences. He works as an associate editor for IET Electronics letters and an academic editor for Biomed research international journal, Hindawi. He is also working as guest editor for a special issue on "machine learning and deep learning for physiological data analysis" for Frontiers in Physiology Journal.